


This supplier portal document specifies requirements & request for external laboratories executing services for Sekisui S-lec BV Netherlands.

 <p>Leading technology, leading the way</p>		Calibration	Product testing	Product inspection
Specific requirements and requests		REQUIRED (●) REQUESTED (○)		
Laboratory scope	ISO/TS requirement 7.6.2.3.	●	●	●
ISO-9001	The laboratory shall be third party registered to ISO 9001 (latest version) by an accredited third-party certification body.	●	●	●
ISO/IEC 17025	The laboratory shall be accredited to ISO/IEC 17025 (latest version) or national equivalent	●	●	●
Additional Important Information				
<p>ISO/TS 7.6.3.2 External/commercial/independent laboratory facilities used for inspection, test or calibration services by the organization shall have a defined laboratory scope that includes the capability to perform the required inspection, test or calibration, and either there shall be evidence that the external laboratory is acceptable to the customer, or the laboratory shall be accredited to ISO/IEC 17025 or national equivalent.</p> <p>Laboratory Scope: A defined scope that includes its capability to perform the required inspection, test or calibration services. This laboratory scope shall be included in the management system documentation. The laboratory shall specify and implement, as a minimum, technical requirements for:</p> <p>1. adequacy of the laboratory procedures 2. competency of the laboratory personnel 3. testing the product 4. capabilities to perform these services correctly, traceable to the relevant standard (such as ASTM, EN, etc.) 5. review of the related documents</p>				